

## Operational Guidelines for the Laboratory Animal Biochemical

### Analysis Service

Article 1 The Operational Guidelines for the Laboratory Animal Biochemical Analysis Service were established to maximize the utility and effectiveness of the device mentioned herein.

Article 2 Details of the device are as follows:

1. Brand and type: IDEXX Chemistry Analyzer (Catalyst One)
2. Performance:
  - (1) Dry-slide technology is used to provide results with optimal accuracy.
  - (2) Reference values for 21 test items and 39 specific species and ages are available.
  - (3) Results of an analyte can be obtained within 8 minutes.
3. IDEXX VetLab™ Station—a test data integration system.

Article 3 The device is location in the Laboratory Animal Center on the first floor of the Back Building of the United Medical Building

Article 4 The main service targets of the device are researchers from Taipei Medical University and its three associated hospitals.

Article 5 Service items of this device consist of 21 types of dry-slide CLIPs:

生化項目	縮寫	全名
白蛋白	ALB	Albumin
鹼磷酶	ALKP	Alkaline Phosphatase
丙胺酸轉胺酶	ALT	Alanine Aminotransferase
胰澱粉酶	AMYL	Amylase
天門冬酸轉胺酶	AST	Aspartate Aminotransferase
血中尿素氮	BUN	Blood Urea Nitrogen
鈣	Ca	Calicum
膽固醇	CHOL	Cholesterol
肌酸酐	CREA	Creatinine
氯離子	Cl	Chloride
果糖胺	FRU	Fructosamine
加瑪麩氨肌轉換酶	GGT	Gamma-glutamyltransferase

血糖	GLU	Glucose
鉀離子	K	Potassium
乳酸脫氫酶	LDH	L-lactate dehydrogenase
鈉離子	Na	Sodium
磷離子	PHOS	Inorganic Phosphate
總膽紅素	TBIL	Total Bilirubin
總蛋白	TP	Total Protein
三酸甘油脂	TRIG	Triglyceride
尿酸	URIC	uric acid

Article 6 **Reservation method:**

1. Please register use of the device in the reservation system of Core Facility Center or Laboratory Animal Center 1–10 days prior to the prescheduled service day. The technician of Laboratory Animal Center can decide whether to provide services according to actual reservation situations.
2. Technical service time: Monday to Friday, 09:00 a.m. to 05:00 p.m.

Article 7 The guidelines for using the device are as follows:

1. **Analysis reports of the service are provided in hard copies.**
2. The IDEXX Catalyst One biochemical analyzer uses 21 types of dry-slide CLIPs, which are disposable after a single use. A box of 12 CLIPs is required for each operation. The Laboratory Animal Center neither safeguards the remaining unused CLIPs for users nor accepts CLIPs prepared by them to control the accuracy of the analysis results. Researchers must consider the sample amount they plan to analyze and make necessary adjustment.
3. The device can accept serum samples (without anticoagulants) or plasma samples (those using the anticoagulant ethylenediaminetetraacetic acid shall be rejected).
4. In principle, the samples should not be hemolyzed. Hemolyzed samples severely affect the testing of AST, ALT, LDH, and TBIL and partially affect other items. Researchers should consider the effect of hemolysis.
5. Fasting of the animal can affect the test results of certain items (i.e., blood glucose, lipid, and inorganic phosphor), which researchers should consider. If fasting of the animals is required, **rats can fast for**

15–20 hours and mice should not fast for longer than 4 hours; water fasting should not be implemented.

6. **Required sample calculations:**

- (1) The required sample volume for analyzing the first item is 60  $\mu\text{L}$ ; increments of 10  $\mu\text{L}$  are required for each item added for analysis.
  - (2) Limited by the size of the device, up to 18 analysis items can be accepted in one analysis. If a 19<sup>th</sup> item must be analyzed, a second analysis should be conducted and the required sample volume recalculated.
  - (3) For example, when 19 items are to be analyzed, the required sample volume for the 1<sup>st</sup> to 18<sup>th</sup> items is 230  $\mu\text{L}$  in the first analysis. An additional 60  $\mu\text{L}$  is required for the 19<sup>th</sup> item in the second analysis. Therefore, the analysis of 13 items requires a total sample volume of 290  $\mu\text{L}$ .
7. If serum or plasma samples cannot be immediately sent for testing after collection, they should be preserved at a low temperature. A general recommendation is that samples can be stored for approximately 1 month at  $-80^{\circ}\text{C}$  and approximately 10 days at  $-20^{\circ}\text{C}$ . Storage at  $4^{\circ}\text{C}$  for a long time may affect the test values of enzymes.

Article 8 The charging standard of this device is determined to ensure optimal service quality and extend the effective service duration of the device. Upon resolution of the Meeting of the Office of Research and Development, individuals or units that use the device must pay for necessary consumables, maintenance, and operator services.

1. Charging standard: NT\$800/box for an analytical item, which can be used for the analysis of 12 samples.
2. Payment method: The Laboratory Animal Center calculates technical service uses and relevant fees every month and sends payment notices to users. Users must pay and verify the technical service fees within 3 months after the payment notice is issued. Users can visit the Cashier Section of Taipei Medical University to pay the fees, or have their expenditures charged from research project funds or school budgets.

Article 9 Device contact person:

Laboratory Animal Center technician: Wang, Chueh-Yi, extension: 7153, 7264

Laboratory Animal Center technician: Wu, Wen-Chi, extension: 7147, 7256

Article 10 These Guidelines shall be implemented upon approval by an Office Meeting of the Office of Research and Development; the same procedure

applies to any amendment.